

1HNMR  
Sample NO-P-76



Current Data Parameters  
NAME Sample NO-P-76  
EXPNO 1  
PROCNO 1

F2 - Acquisition Parameters  
Date\_ 20250930  
Time 10.54 h  
INSTRUM Avance  
PROBHD Z163739\_1083  
PULPROG zg30  
TD 65536  
SOLVENT CDC13  
NS 16  
DS 2  
SWH 8196.721 Hz  
FIDRES 0.250144 Hz  
AQ 3.9976960 sec  
RG 32  
DW 61.000 usec  
DE 13.97 usec  
TE 298.2 K  
D1 1.00000000 sec  
TD0 1  
SF01 400.5024731 MHz  
NUC1 1H  
P0 2.50 usec  
P1 7.50 usec  
PLW1 25.48000000 W

F2 - Processing parameters  
SI 65536  
SF 400.4999950 MHz  
WDW EM  
SSB 0  
LB 0.30 Hz  
GB 0  
PC 1.00

